

ASO Immunoturbidity Method



Intended Use: Kit for the quantitative determination of ASO in Human Serum and Plasma.

Pack Size: 50 ml

Contents:

Reagent-1

ASO Buffer Reagent : 1 x 40 ml

: 1 x 10 ml

Reagent-2

ASO Latex Reagent

Reagent-3

ASO Calibrator : 1 x 0.5 ml Product Insert : 01 No.

Directions for Use:

Use the process instruction outlined in the kit insert

- For in-vitro diagnostics use only [VD]
- Do not inhale the reagent or pipette by mouth.
- Avoid contact with skin/eyes. In case of contact, wash off immediately with plenty of water and seek medical attention as early as possible.
- 4. Do not use if package is damaged.
- Dispose waste product in accordance with local Govt. regulations.

Storage and Stability:

The unopened reagent is stable up to the expiry date mentioned on the label when stored between 2 - 8°C.

MLC.No.: MFG/IVD/2021/000002

LOT

M.R.P. (Rs.) : (Incl. of all taxes)

Manufactured in India by: Vanguard Diagnostics Private Limited C-123, Phase-I, Okhla Industrial Area New Delhi-110 020, INDIA, Tel:011-4172 7222 Email:info@vanguarddiagnostics.com

Ready to use.

ASO BUFFER Immunoturbidity Method

(Reagent 1)

For in vitro Diagnostic use only [IVD]. MLC.No.: MFG/IVD/2021/000002

STORE AT 2 - 8°C. 2°C Manufactured in India by

Vanguard Diagnostics Private Limited C-123, Phase-I, Okhla Industrial Area New Delhi-110 020, INDIA. Tel:011-4172 7222 Email:info@vanguarddiagnostics.com

-8°C

LOT :



ASO LATEX Immunoturbidity Method (Reagent 2)

CE

STORE AT 2 - 8°C. 2°C

For in vitro Diagnostic use only [IVD] MLC.No.: MFG/IVD/2021/000002

Manufactured in India by: Vanguard Diagnostics Private Limited C-123, Phase-I, Okhla Industrial Area New Delhi-110 020, INDIA. Tel:011-4172 7222 Email:info@vanguarddiagnostics.com

g rot ₩



PM30008L702

 (ϵ)

40 ml