



# CREATININE

## (Modified Jaffe's) Liquid Reagent



**Intended Use :** Kit for the quantitative determination of Creatinine in human serum and plasma.

### Contents:

**Pack Size : 4 x 100 ml**

**▽ : 400 Test**

#### Reagent-1

Creatinine Reagent : 4 x 100 ml

#### Reagent-2

Standard : 1 x 10 ml

Insert : 01 No.

#### Directions for Use:

Use the process instruction outlined in the kit insert

#### Precautions:

1. For *in-vitro* diagnostics use only **IVD**.
2. Do not inhale the reagent or pipette by mouth.
3. Avoid contact with skin/eyes. In case of contact, wash off immediately with plenty of water and seek medical attention as early as possible.
4. Do not use if package is damaged.
5. Dispose waste product in accordance with local Govt. regulations.

#### Storage and Stability:



The unopened reagent is stable up to the expiry date mentioned on the label when stored between 2 - 8°C.

**MLC.No. : S(0079)/15/MB**

**LOT** :



**M.R.P. Rs.:**  
(Incl. of all taxes)



**Manufactured by:**  
**Vanguard Diagnostics Private Limited**  
C-123, Phase-I, Okhla Industrial Area  
New Delhi-110 020, INDIA. Tel:011-4172 7222  
Email:info@vanguarddiagnostics.com

PM30008L49



# CREATININE

## Liquid Reagent (Reagent 1)



**100 ml**

Ready to use

For *in vitro* Diagnostic use only **IVD**.

**STORE AT 2 - 8°C.**

**MLC.No. : S(0079)/15/MB**

**Manufactured by:**  
**Vanguard Diagnostics Private Limited**  
C-123, Phase-I, Okhla Industrial Area  
New Delhi-110 020, INDIA. Tel:011-4172 7222  
Email:info@vanguarddiagnostics.com

PM30008L50

**LOT** :



# CREATININE

## Liquid Reagent (Reagent 2)



**10 ml**

**STANDARD**

Ready to use

**STORE AT 2 - 8°C.**

For *in vitro* Diagnostic use only **IVD**.  
**MLC.No. : S(0079)/15/MB**

**Manufactured by:**  
**Vanguard Diagnostics Private Limited**  
C-123, Phase-I, Okhla Industrial Area  
New Delhi-110 020, INDIA.  
Tel:011-4172 7222  
Email:info@vanguarddiagnostics.com

PM30008L51

**LOT** :

