



# HDL CHOLESTEROL

(Direct Reagent Kit) Liquid Reagent



**Intended Use :** Kit for the quantitative determination of HDL Cholesterol in human serum and plasma.

**Contents:**

- Reagent 1 :** R1 Reagent : 2 x 30 ml  
**Reagent 2:** R2 Reagent : 2 x 10 ml  
**Insert** : 01 No.

**Pack Size :** 2 x 40 ml  
 ▽ : 160 Test

**Storage and Stability:** 2°C - 8°C

The unopened reagent is stable up to the expiry date mentioned on the label when stored between 2 - 8°C.

**Directions for Use:**

Use the process instruction outlined in the kit insert

**Precautions:**

- For *in-vitro* diagnostics use only **IVD**.
- Do not inhale the reagent or pipette by mouth.
- Avoid contact with skin/eyes. In case of contact, wash off immediately with plenty of water and seek medical attention as early as possible.
- Do not use if package is damaged.
- Dispose waste product in accordance with local Govt. regulations.

**MLC.No. :** S(0079)/15/MB

**LOT** :

:

:

**M.R.P. Rs.:**  
 (Incl. of all taxes)

**Manufactured by:**  
**Vanguard Diagnostics Private Limited**  
 C-123, Phase-I, Okhla Industrial Area  
 New Delhi-110 020, INDIA. Tel:011-4172 7222  
 Email:info@vanguarddiagnostics.com

PM30008L60



# HDL CHOLESTEROL

(Direct Reagent Kit) Liquid Reagent  
 (Reagent 1)



Ready to use

For *in vitro* Diagnostic use only **IVD**.

**STORE AT** 2 - 8°C. 2°C - 8°C

**MLC.No. :** S(0079)/15/MB

30 ml

**Manufactured by:**  
**Vanguard Diagnostics Private Limited**  
 C-123, Phase-I, Okhla Industrial Area  
 New Delhi-110 020, INDIA. Tel:011-4172 7222  
 Email:info@vanguarddiagnostics.com

PM30008L61  
**LOT** :  
 :  
 :



# HDL CHOLESTEROL

(Direct Reagent Kit) Liquid Reagent  
 (Reagent 2)



Ready to Use

For *in vitro* Diagnostic use only **IVD**.

**STORE AT** 2 - 8°C. 2°C - 8°C

**MLC.No. :** S(0079)/15/MB

10 ml

**Manufactured by:**  
**Vanguard Diagnostics Private Limited**  
 C-123, Phase-I, Okhla Industrial Area  
 New Delhi-110 020, INDIA.  
 Tel:011-4172 7222  
 Email:info@vanguarddiagnostics.com

PM30008L62  
**LOT** :  
 :  
 :